



EC Certificate Production Full Quality Assurance System: Certificate GB19/964561

The management system of

# LaproSurge Limited

5b Fishers Industrial Estate, Wiggshall Road,  
Watford, Hertfordshire, WD18 0FN, UK

has been assessed and certified as meeting the requirements of

## Directive 93/42/EEC

on medical devices, Annex V

**Restricted to the aspects of manufacture concerned with securing and maintaining sterile conditions**

For the following products

- Sterile smoke venting filter
- Sterile laparoscopic camera sleeves
- Sterile disposable monopolar cables

Where the above scope includes class IIb or class III medical device(s), a valid EC Type Examination Certificate according to Annex III is a mandatory requirement for each device in addition to this certificate to place that device on the market.

This certificate is valid from 01 April 2021 until 24 May 2024 and remains valid subject to satisfactory surveillance audits.

Issue 5. Certified since 01 October 1999.

Certification is based on reports numbered GB/PC/ 06917

Authorised by

Global Medical Devices Head of Notified Body

### SGS Belgium NV, Notified Body 1639

SGS House Noorderlaan 87 2030 Antwerp Belgium  
t +32 (0)3 545-48-48 f +32 (0)3 545-48-49 www.sgs.com

LPMD5008 - Certificate CE1639 Annex V, EN Rev. 02

Page 1 of 1



This document is a Web version of SGS certificate for electronic use exclusively. It shall only be available by clicking on SGS Certification Mark which has been posted on Your website. It shall not be printed in anyway. This document is copyright protected. No content or appearance may be reproduced without the express written permission of SGS. Any misuse, alteration, forgery or falsification is unlawful.